

Declaration of Conformity

According to EN45014

Manufacturer's Name: Webscan, Inc.
Manufacturer's Address: 1254 Sherman Drive
 Unit #1
 Longmont, CO 80501



Declares that the CE-marked product:

Item	Model Number	Manufacturer
TruCheck USB Bar Code Verifier	TC-833 (Gray), TC-866 (Axicon Blue), TC-830 (Gray), TC-865 (Axicon Blue), TC-836 (Gray), TC-867 (Axicon Blue)	Webscan, Inc.

Complies with:

The essential safety requirements of the Low Voltage Directive 2006/95/EC and the EMC Directive 2004/108/EC. In addition, said products also comply with the applicable requirements of:

Immunity Requirement	Specification	Test Method	Performance Criteria
EN 61326-1:2006 Electrical Equipment for Measurement, Control and Laboratory Use	Electrostatic discharge	IEC 61000-4-2: 2001-04	(A) No Degradation
	Radiated RF Immunity	IEC 61000-4-3:2006-02	(A) No Degradation
	Electrical Fast Transients / Burst	IEC 6100-4-4: 2004-07	(A) No Degradation
	Surge Immunity	IEC 61000-4-5: 2005-11	(A) No Degradation
	Conducted RF Immunity	IEC 61000-4-6:2006-05	(A) No Degradation
	Voltage Dips, Interrupts	IEC 61000-4-11:2004-03	(A) No Degradation for parts A through D (C) Recovered with Operator Intervention for parts E and F
Emissions Requirement			
Limits and methods of measurement of radio disturbance characteristics of industrial, scientific and medical (ISM) radio-frequency equipment	Radiated and Conducted Emissions	EN55011:2009 +A1:2010 Class B Group 1	Passing under specified limits

Safety Requirement	
Directive	Standard
European Low Voltage Directive (LVD) 2006/95/EC	EN 60950-1 Safety Requirements for Information Technology Equipment

Glenn Spitz,
 President, Webscan Inc.
 2nd December 2011